



Schofield Waterworks  
200 Park Street  
Schofield, WI 54476

Project: Quarterly PFAS  
Project Number: 2023 WDNR Drinking Water Requirements  
Project Manager: Mark Thuot

Reported:  
6/22/23 9:28

Work Order:  
CB05912

**Sample Results**

**Sample: EP3 (PFAS)**

**CB05912-01 (DW) Sampled: 06/05/23 10:15**

Analyte	Result	Qualifier	LOD	LOQ	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
<b>Semi-Volatiles</b>											
11-chloroicosafafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		0.32	1.0		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	ND		0.35	1.1		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.38	1.2		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.42	1.4		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
N-ethyl perfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		0.48	1.6		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
n-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		0.41	1.3		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorobutanesulfonic acid (PFBS)	0.63	J	0.31	1.0		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorodecanoic acid (PFDA)	ND		0.34	1.1		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorododecanoic acid (PFDoA)	ND		0.23	0.79		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluoroheptanoic acid (PFHpA)	ND		0.45	1.5		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanoic acid (PFHxA)	0.52	J	0.48	1.6		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanesulfonic acid (PFHxS)	3.7		0.35	1.1		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorononanoic acid (PFNA)	ND		0.47	1.5		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanoic acid (PFOA)	1.1	J	0.50	1.6		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanesulfonic acid (PFOS)	1.1		0.32	1.0		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorotetradecanoic acid (PFTA)	ND		0.35	1.1		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluorotridecanoic acid (PFTrDA)	ND		0.44	1.4		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
perfluoroundecanoic acid (PFUnA)	ND		0.31	1.0		ng/L	6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFHxA	89%		Limits: 70-130%				6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-HFPODA	87%		Limits: 70-130%				6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFDA	95%		Limits: 70-130%				6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) d5-NEtFOSAA	83%		Limits: 70-130%				6/7/23 7:05	6/8/23 1:06	RAW	EPA 537.1, Rev 2.0	2



Schofield Waterworks  
200 Park Street  
Schofield, WI 54476

Project: Quarterly PFAS  
Project Number: 2023 WDNR Drinking Water Requirements  
Project Manager: Mark Thuot

Reported:  
6/22/23 9:28

Work Order:  
CB05912

**Sample: EP3 Field Blank**

**CB05912-02 (DW) Sampled: 06/05/23 10:15**

Analyte	Result	Qualifier	LOD	LOQ	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
<b>Semi-Volatiles</b>											
11-chloroicosafafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		0.32	1.0		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	ND		0.35	1.1		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.38	1.2		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.42	1.4		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
N-ethyl perfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		0.48	1.6		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
n-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		0.41	1.3		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorobutanesulfonic acid (PFBS)	ND		0.31	1.0		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorodecanoic acid (PFDA)	ND		0.34	1.1		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorododecanoic acid (PFDoA)	ND		0.23	0.79		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluoroheptanoic acid (PFHpA)	ND		0.45	1.5		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanoic acid (PFHxA)	ND		0.48	1.6		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanesulfonic acid (PFHxS)	ND		0.35	1.1		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorononanoic acid (PFNA)	ND		0.47	1.5		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanoic acid (PFOA)	ND		0.50	1.6		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanesulfonic acid (PFOS)	ND		0.32	1.0		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorotetradecanoic acid (PFTA)	ND		0.35	1.1		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluorotridecanoic acid (PFTTrDA)	ND		0.44	1.4		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
perfluoroundecanoic acid (PFUnA)	ND		0.31	1.0		ng/L	6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFHxA	88%		Limits: 70-130%				6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-HFPODA	87%		Limits: 70-130%				6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFDA	94%		Limits: 70-130%				6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) d5-NEtFOSAA	89%		Limits: 70-130%				6/12/23 6:15	6/12/23 17:52	RAW	EPA 537.1, Rev 2.0	2